

Gas Chromatography Mass Spectrometry

TROUBLESHOOTING GUIDE



SHIMADZU

Excellence in Science



Contents

Introduction	04
No Peaks	05
Broad Peaks	06
High Baseline (Column Bleed)	07
Unstable Baseline (Spiking, Noise, Drift)	08
Tailing Peaks	09
Changes in Response	10
Carryover/Ghost Peaks	11
Fronting Peaks	12
Poor Peak Resolution	13
Split Peaks	14
Retention Time Variability	15
Changes in MS Resolution	16
Undesired Fragmentation	17



Introduction

Instrument downtime is often costly and time consuming, but frequently the problems can be resolved quickly with some troubleshooting knowledge.

This Gas Chromatography Mass Spectrometry Troubleshooting Guide is designed to assist chromatographers assess common GCMS problems. Shimadzu have included how to effectively troubleshoot and fix these issues to allow you to get your system back up and running and continue your analyses.

Basic Steps

Follow these three steps to isolate the source of the problems. Check the obvious causes first and change only one thing at a time!

Check the Basics:

- Power supply
- Electrical connections
- Signal connections
- Syringe condition
- Sample preparation
- Analytical conditions
- Temperature settings
- Gas purity
- Gas flows
- MS vacuum
- Collision gas (level & pressure)
- Rough pump (oil level & gas ballast)

Identify the Cause:

- Define the problem clearly; for example, "Over the last four days, only the phenols in my sample have been tailing."
- Review sample and maintenance records to identify trends in the data or problem indicators, such as area counts decreasing over time or inlet maintenance not being performed as scheduled.
- Use a logical sequence of steps to isolate possible causes.

Document Everything:


- Document all troubleshooting steps and results; this may help you identify and solve the next problem faster.
- Always inject a test mix and compare to previous data to ensure restored performance.

Still having problems?

Still struggling? Let us know!!!

gc_gcms_support@shimadzu.co.uk

No Peaks




Causes

Solutions

Injection problems	<ul style="list-style-type: none"> - Blocked syringe; clean or replace syringe. - Verify there is sample in the vial. - Injecting into wrong inlet; reset autosampler. - Verify carrier gas is flowing.
Broken column	<ul style="list-style-type: none"> - Replace column.
Column installed into wrong inlet or detector	<ul style="list-style-type: none"> - Reinstall column.
Detector problems	<ul style="list-style-type: none"> - Ensure MS is an appropriate technique for the analyte(s) - Check the vacuum is sufficient and stable, perform a leak check. - Perform and check the MS tune. - Check the ion source temperatures and gas flows are as expected and stable. - Check the collision gas pressure is correct and stable. - If the method contains scheduled events, such as MRM, ensure the analyte elutes within the correct event window. - Check that the concentration of the analyte is above the limit of detection. - Blown filament, switch to second filament and replace the first as soon as practicable.


Broad Peaks



Causes	Solutions
High dead volume	- Minimise dead volume in the GC system; verify proper column installation, proper connectors, proper liners, etc.
Low flow rates	- Verify carrier gas flow rate and adjust if needed.
Slow GC oven program	- Increase GC oven programming rate.
Poor analyte / solvent focusing	- Lower GC oven start temperature.
Column film is too thick	- Reduce retention of compounds by decreasing film thickness and length.
Sample carryover	- See Carryover/Ghost Peaks solutions.

High Baseline




(Column Bleed)



Causes	Solutions
Improper column conditioning	- Increase conditioning time and/or temperature.
Contamination	- Trim column and/or heat to maximum temperature to remove contaminants. - Replace carrier gas and/or detector gas filters. - Clean injector and ion source.
Leak in system causing oxidation of stationary phase	- Check for oxygen leaks across the entire system and replace seals and/or filters. Perform a manual tune and look for the presence of m/z 18, 28 & 32. Large peaks will indicate a leak. - Check the column fittings on the inlet and MS transfer line. - Replace column. - Check the seal around the access door to the MS.


Unstable Baseline

(Spiking, Noise, Drift)




Causes	Solutions
Carrier gas leak or contamination	<ul style="list-style-type: none"> - Leak check the system and check that the flow and pressure controllers are functioning correctly. - Replace carrier gas and/or detector gas filters.
Inlet or detector contamination	<ul style="list-style-type: none"> - Clean the inlet and ion source. Schedule and perform regular maintenance.
Column contamination or stationary phase bleed	<ul style="list-style-type: none"> - Condition, trim, and rinse column.
Septum coring / bleed	<ul style="list-style-type: none"> - Replace septum. - Inspect inlet liner for septa particles and replace if needed. - Replace the solvent in the wash vials. - Perform a manual tune, the presence of m/z 73, 147, 207, 281 & 355 will indicate siloxane contamination from damaged septa.
Aging filament	<ul style="list-style-type: none"> - Switch to second filament and replace the first as soon as practicable.
Variable carrier gas or detector gas flows	<ul style="list-style-type: none"> - Leak check the system and check that the flow and pressure controllers are functioning correctly. - Check collision gas pressure.
Detector not ready	<ul style="list-style-type: none"> - Allow enough time for ion source and transfer line temperatures to equilibrate.

Tailing Peaks




Causes	Solutions
Adsorption due to surface activity or contamination	<ul style="list-style-type: none"> - Use properly cleaned and deactivated liner and column. - Trim inlet end of column. - Replace column if damaged.
Adsorption due to chemical composition of compound	<ul style="list-style-type: none"> - Derivatise compound.
Leak in system	<ul style="list-style-type: none"> - Check for leaks at all connections, replace critical seals if needed.
Column installation issues	<ul style="list-style-type: none"> - Minimise dead volume. - Verify correct installation depth. - Verify that the column is cut properly (square).

Changes in Response




Causes	Solutions
Sample issues	<ul style="list-style-type: none"> - Check sample concentration. - Check sample preparation procedure. - Check sample decomposition/shelf life.
Syringe problems	<ul style="list-style-type: none"> - Replace syringe. - Check autosampler operation.
Electronics	<ul style="list-style-type: none"> - Verify signal settings and adjust if needed.
Dirty or damaged detector	<ul style="list-style-type: none"> - Clean the ion source and replace the electron multiplier if required. - Check and replace filaments if required.
Flow / temperature settings wrong or variable	<ul style="list-style-type: none"> - Verify steady flow rates and temperatures, then adjust settings and/or replace parts if needed.
Adsorption / reactivity	<ul style="list-style-type: none"> - Remove contamination and use properly deactivated liner and column.
Leaks	<ul style="list-style-type: none"> - Check for leaks at all connections and repair connections as needed.
Change in sample introduction / injection method	<ul style="list-style-type: none"> - Verify injection technique and change back to original technique. - Check that split ratio is correct. - Verify that the splitless hold time is correct.

Carryover/Ghost Peaks




Injection 1



Injection 2


Causes	Solutions
Contaminated syringe or rinse solvent	<ul style="list-style-type: none"> - Replace rinse solvent. - Rinse or replace syringe.
Backflash (sample volume exceeds liner volume)	<ul style="list-style-type: none"> - Inject a smaller amount. - Use a liner with a large internal diameter. - Increase head pressure (i.e. flowrate) to contain the vapour cloud. - Use slower injection rate. - Lower inlet temperature. - Use liner with packing. - Use pressure-pulse injection. - Use online calculator to check expansion volume.
Last analysis ended too soon	<ul style="list-style-type: none"> - Extend analysis time to allow all components and/or matrix interferences to elute.

Fronting Peaks




Causes	Solutions
Incompatible stationary phase	- Choose appropriate stationary phase.
Column overloading	- Reduce amount injected, dilute sample or increase split ratio. - Increase column inner diameter and/or film thickness.

Poor Peak Resolution




Causes	Solutions
Non-selective stationary phase	- Choose an appropriate stationary phase and column dimensions.
Poor efficiency	- Optimise carrier gas linear velocity and GC oven temperature program.
Sample overload	- Adjust sample concentration or amount on column by increasing split ratio.
Incorrect analytical conditions used	- Verify temperature program, flow rates, and column parameters.
MS method settings	- Use different m/z when using selected ion monitoring (SIM) or a different transition in multiple reaction monitoring (MRM). - Set quadrupole to 'High' resolution for analytes affected in MRM mode.

Split Peaks





Causes	Solutions
Mismatched solvent / stationary phase polarity	- Adjust solvent or stationary phase to allow wetting.
Incomplete vaporisation	- Add surface area, such as wool, to the inlet liner to enhance vaporisation. - Use proper inlet temperature.
Sample loading capacity exceeded	- Inject less sample (dilute, use split injection, reduce injection volume).
Fast autosampler injection into open liner	- Use wool or slow injection speed.
Detector saturation	- Reduce the injection volume or inject a lower concentration of sample. - Set quadrupoles to 'High' resolution for analytes affected in MRM mode.

Retention Time Variability



Causes	Solutions
Leaks	- Leak check inlet and any column connections. - Replace septa, O-rings, etc.
Analyte adsorption	- Maintain inlet liner and GC column. - Use properly deactivated liners and columns.
Resolution / integration issues	- Avoid sample overload by diluting sample or increasing split ratio.
Incorrect column / oven temperature program	- Verify column temperature and oven temperature program.
Incorrect or variable carrier gas linear velocity	- Verify the carrier gas linear velocity. - Repair or replace parts if necessary.
Poor control of oven temperature programming	- Confirm GC oven program falls within instrument specifications.
Incorrect oven equilibration time	- Extend GC oven equilibration time.
If manual injection, inconsistencies between pushing start and injection procedure	- Use autosampler or standardise manual injection procedure.

Changes in MS Resolution





Causes

MS out of tune

Solutions

- Perform and check tune

Undesired Fragmentation



Causes

Ion source setting too harsh

Collision energy too low / high

Aging filament

Contaminated source


Solutions

- Check source temperatures are appropriate for the analyte.
- Check ionisation voltage is appropriate for the analyte.

- Check and optimise the collision cell gas pressure and collision energy.

- Change to the second filament and replace the first as soon as practicable.

- Clean the ion source and perform a tune.





www.shimadzu.co.uk
info@shimadzu.co.uk | 01908 55 22 09

Technical Support:
gc_gcms_support@shimadzu.co.uk